

Republic of the Philippines Department of Health

ADELA SERRA TY MEMORIAL MEDICAL CENTER

Capitol Hills, Tandag City, Surigao del Sur Telefax No. 086-211-4306, E-mail add: astmmc hbac@yahoo.com.ph



INVITATION TO BID FOR PROCUREMENT OF VARIOUS MEDICAL EQUIPMENT

IB No. 2020-12- 14 (17)

- 1. The Adela Serra Ty Memorial Medical Center, through HFEP CY 2019 intends to apply the sum of Four Million Seven Hundred Thousand Pesos Only (Php 4,700,000.00) being the Approved Budget for the Contract (ABC) to payments under the contract of Procurement of Various Medical Equipment. Bids received in excess of the ABC shall be automatically rejected at bid opening
- 2. The Adela Serra Ty Memorial Medical Center now invites bids for Procurement of Various Medical Equipment. Delivery of the Goods required is 15-30 Calendar days. Bidders should have completed, within 5 years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instructions to Bidders.
- 3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary "pass/fail" criterion as specified in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act (RA) 9184, otherwise known as the "Government Procurement Reform Act".
 - (i) Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183.
- 4. Interested bidders may obtain further information from *Adela Serra Ty Memorial Medical Center* and inspect the Bidding Documents at the address given below during *8:00am to 12:00; 1:00 to 5:00pm, Monday thru Friday.*
- 5. A complete set of Bidding Documents may be purchased by interested Bidders on [November 21, 2020 to December 14, 2020] from the address below and upon payment of a nonrefundable fee for the Bidding Documents in the amount of Five Thousand Pesos Only (Php 5,000.00). It may also be downloaded free of charge from the website of the Philippine Government Electronic Procurement System (PhilGEPS) and the website of the Procuring Entity, provided that Bidders shall pay

the applicable fee for the Bidding Documents not later than the submission of their bids.

- The Adela Serra Ty Memorial Medical Centerwill hold a Pre-Bid Conference on [10:00am; December 01, 2020] at 3rd Floor, AVR, OPD Building, ASTMMC, Tandag City, which shall beopen only to all interested parties who have purchased the Bidding Documents.
- 7. Bids must be delivered to the address below on or before [10:00AM, December 08, 2020]. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 18. Bid opening shall be on [10:01am; December 14, 2020] 3rd Floor, AVR, OPD Building, ASTMMC, Tandag City. Bids will be opened in the presence of the Bidders' representatives who choose to attend at the address below. Late bids shall not be accepted.
- The Adela Serra Ty Memorial Medical Center reserves the right to accept or reject any bid, to annul the bidding process, and to reject all bids at any time prior to contract award, without thereby incurring any liability to the affected bidder or bidders.
- 9. For further information, please refer to:

VOLTAIRE S. EGNORA, MD, FPCP HBAC Office ASTMMC Tel No. 086-211-4306 astmmc_hbac@yahoo.com.ph

> VOLTAIRE S. EGNORA, MD, FPCP [BAC Chairman]

Bid Data Sheet

ITB Clause	

1.1	The Procuring Entity is Adela Serra Ty Memorial Medical Center	
	The name of the Contract is Procurement of Various Medical Equipment.	
	The identification number of the Contract is IB No. 2020-12-14 (17)	
2	The Funding Source is:	
	The Government of the Philippines (GOP) through <i>HFEP CY 2019</i> in the amount of <i>FOUR MILLION SEVEN HUNDRED PESOS ONLY (PHP 4,700,000.00)</i>	
	The name of the Project is: Procurement of Various Medical Equipment.	
3.1	No further instructions.	
5.1	No further instructions.	
5.2	None of the circumstances mentioned in the ITB Clause exists in this Project. Foreign bidders, except those falling under ITB Clause Error! Reference source not found. , may not participate in this Project	
5.4	The Bidder must have completed, within the period specified in the Invitation to Bid and ITB Clause 12.1(a)(ii), a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.	
	For this purpose, similar contracts shall refer to Procurement of Various Medical Equipment.	
5.5	NFCC = [(Current assets minus current liabilities) (15)] minus the value of all outstanding or uncompleted portions of the projects under ongoing contracts, including awarded contracts yet to be started, coinciding with the contract to be bid.	
	Note: NFCC should be notarized	
7	No further instructions.	
8.1	"Subcontracting is not allowed."	
8.2	"Not applicable"	
9.1	The Procuring Entity will hold a Pre-bid conference for this Project on [10:00am; December 01, 2020] at [3rd Floor, AVR, OPD Building, ASTMMC, Tandag City].	
10.1	The Procuring Entity's address is:	

	[Adela Serra Ty Memorial Medical Center, Capitol Hills, Tandag City, Surigao del Sur]	
	VOLTAIRE S. EGNORA, MD, FPCP	
	HBAC- Chairman	
	[Tel No. 086-2114306/Fax No. 086-2114306]	
	[astmmc_hbac@yahoo.com.ph)]	
12	Additional Requirements	
	Print out of Bidding Documents to be attached after Omnibus Sworn Statement.	
12.1 (a)(ii)	The statement of all ongoing and completed government and private contracts shall include all such contracts within [5 years] prior to the deadline for the submission and receipt of bids.	
13.1	"No additional requirements."	
13.1(b)	No further instructions.	
13.1(c)	"No additional requirements."	
13.2	The ABC is FOUR MILLION SEVEN HUNDRED THOUSAND PESOS ONLY (PHP 4,700,000.00). Any bid with a financial component exceeding this amount shall not be accepted.	
15.4(a)(iv)	"No incidental services are required."	
15.4(b)	"Not applicable"	
16.1(b)	The Bid prices for Goods supplied from outside of the Philippines shall be quoted in Philippine Pesos.	
16.3	"Not applicable"	
17.1	Bids will be valid until [120 calendar days].	
	From the date of bid opening.	
18.1	The bid security shall be in the form of a Bid Securing Declaration , or any of the following forms and amounts:	
	1. The amount of not less than <i>Php 94,000.00</i> , if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or	
	2. The amount of not less than <i>Php 235,000.00</i> if bid security is in Surety	

	Bond.		
18.2	The bid security shall be valid until [120 Calendar Days] from the opening of bids.		
20.3	 Each Bidder shall submit the following: i. Eligibility and Technical Documents, [1] original and [3] copies, sealed into First envelop. ii. Financial Documents, [1] original and [3] copies, sealed into Second envelop. iii. First and Second envelop sealed into Mother Envelop. 		
20.4	All envelopes shall:		
	(a) contain the name of the contract to be bid in capital letters;		
	(b) bear the name and address of the Bidder in capital letters;		
	(c) be addressed to the Procuring Entity's BAC in accordance with ITB Clause 20.1;		
	(d) bear the specific identification of this bidding process indicated in the ITB Clause 1.2; and		
	(e) bear a warning "DO NOT OPEN BEFORE" the date and time for the opening of bids, in accordance with ITB Clause Error! Reference source not found		
21	The address for submission of bids is [Adela Serra Ty Memorial Medical Center, Tandag City, SDS].		
	The deadline for submission of bids is [10:00am, December 14, 2020].		
Error! Reference	The place of bid opening is [3 rd Floor, AVR, OPD Building, ASTMMC, TandagCity].		
source not found.	The date and time of bid opening is [10:01am, December 14, 2020].		
24.2	No further instructions.		
24.3	No further instructions.		
25.1	Members of the BAC, including its staff and personnel, as well as its Secretariat and TWG, are prohibited from making or accepting any kind of communication with any bidder regarding the evaluation of their bids until the issuance of the Notice of Award, unless otherwise allowed in the BDS or in the case of ITB Clause 26.		

Emmore!	No fruthon instructions				
Error!	No further instructions.				
Reference					
source not					
found.					
Error!	Each item to be evaluated and compared with other Bids separately and				
Reference	recommended for contract award separately.				
source not					
found. (a)					
Error!	No further instructions.				
Reference					
source not					
found.					
29.2	List licenses and permits relevant to the Project and the corresponding law requiring it, or state "No additional requirement."				
Error!	Bidders must submit Manually/Electronic filed tax returns.				
Reference					
source not	NOTE: The latest income and business tax returns are those within the last				
found.	six months preceding the date of bid submission.				
30.1	Notwithstanding the eligibility or post-qualification of a bidder, the				
	Procuring Entity concerned reserves the right to review its				
	qualifications at any stage of the procurement process if it has				
	reasonable grounds to believe that a misrepresentation has been				
	made by the said bidder, or that there has been a change in the Bidder's capability to undertake the project from the time it submitted its eligibility requirements. Should such review uncover				
	any misrepresentation made in the eligibility and bidding				
	requirements, statements or documents, or any changes in the				
	situation of the Bidder which will affect its capability to undertake				
	the project so that it fails the preset eligibility or bid evaluation				
	criteria, the Procuring Entity shall consider the said Bidder as				
	ineligible and shall disqualify it from submitting a bid or from				
	obtaining an award or contract.				
30.2	Based on the following grounds, the Procuring Entity reserves the				
	right to reject any and all bids, declare a Failure of Bidding at any				
	time prior to the contract award, or not to award the contract,				
	without thereby incurring any liability, and make no assurance that				
	, , , , ,				
	a contract shall be entered into as a result of the bidding:				
	1.1. If there is <i>prima facie</i> evidence of collusion between				
	, ,				
	appropriate public officers or employees of the				
	Procuring Entity, or between the BAC and any of the				
	bidders, or if the collusion is between or among the				
	bidders themselves, or between a bidder and a third				
	party, including any act which restricts, suppresses				
	or nullifies or tends to restrict, suppress or nullify				
L	in the state of th				

	competition;
	1.2. If the Procuring Entity's BAC is found to have failed in following the prescribed bidding procedures; or
	1.3. For any justifiable and reasonable ground where the award of the contract will not redound to the benefit of the GOP as follows:
	1.4. If the physical and economic conditions have significantly changed so as to render the project no longer economically, financially or technically feasible as determined by the head of the procuring entity;
	1.5. If the project is no longer necessary as determined by the head of the procuring entity; and
	1.6. If the source of funds for the project has been withheld or reduced through no fault of the Procuring Entity.
32.4(f)	List additional contract documents relevant to the Project that may be required by existing laws and/or the Procuring Entity, or state "No additional requirement."

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Lot Number	Description	Quantity	Total	Delivered, Weeks/Months
1	ULTRASONIC SURGICAL & ELECTROSURGICAL SYSTEM COMBINED IN 1 UNIT	1	3,500,000.00	15-30 calendar days
2	STRETCHER HYDRAULIC TYPE	3	750,000.00	15-30 calendar days

3	LARYNGOSCOPE, ADULT	7	262,500.00	15-30 calendar days
4	LARYNGOSCOPE SET FOR PEDIA	5	187,500.00	15-30 calendar days

Bid Form

Date:	
Invitation to Bid ¹ N ^o :	

To: [name and address of Procuring Entity]

Gentlemen and/or Ladies:

Having examined the Bidding Documents including Bid Bulletin Numbers [insert numbers], the receipt of which is hereby duly acknowledged, we, the undersigned, offer to [supply/deliver/perform] [description of the Goods] in conformity with the said Bidding Documents for the sum of [total Bid amount in words and figures] or such other sums as may

¹ If ADB, JICA and WB funded projects, use IFB.

be ascertained in accordance with the Schedule of Prices attached herewith and made part of this Bid.

We undertake, if our Bid is accepted, to deliver the goods in accordance with the delivery schedule specified in the Schedule of Requirements.

If our Bid is accepted, we undertake to provide a performance security in the form, amounts, and within the times specified in the Bidding Documents.

We agree to abide by this Bid for the Bid Validity Period specified in <u>BDS</u> provision for **ITB** Clause **Error! Reference source not found.** and it shall remain binding upon us and may be accepted at any time before the expiration of that period.

Commissions or gratuities, if any, paid or to be paid by us to agents relating to this Bid, and to contract execution if we are awarded the contract, are listed below:²

Name and address	Amount and	Purpose of
of agent	Currency	Commission or gratuity
_	•	
(if none, state "None	")	

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof and your Notice of Award, shall be binding upon us.

We understand that you are not bound to accept the Lowest Calculated Bid or any Bid you may receive.

We certify/confirm that we comply with the eligibility requirements as per **ITB** Clause **Error! Reference source not found.** of the Bidding Documents.

We likewise certify/confirm that the undersigned, [for sole proprietorships, insert: as the owner and sole proprietor or authorized representative of <u>Name of Bidder</u>, has the full power and authority to participate, submit the bid, and to sign and execute the ensuing contract, on the latter's behalf for the <u>Name of Project</u> of the <u>Name of the Procuring Entity</u>] [for partnerships, corporations, cooperatives, or joint ventures, insert: is granted full power and authority by the <u>Name of Bidder</u>, to participate, submit the bid, and to sign and execute the ensuing contract on the latter's behalf for <u>Name of Project</u> of the <u>Name of the Procuring</u> Entity].

We acknowledge that father attached Schedule of Prices	•	y page of this Bid Form, including ejection of our bid.
Dated this	day of	. 20

² Applicable only if the Funding Source is the ADB, JICA or WB.

[signature]	[in the capacity of]
Duly authorized to sign Bid for and on behal	lf of

For Goods Offered From Abroad

Name of Bidder	. Invitation to Bid ³ Number	Page _	of

1	2	3	4		5	6	7	8	9
Lot	Description	Country	Quantity	Units	Unit price CIF	Total	Unit Price	Unit price	Total
No.		of origin			port of entry	CIF or	Delivered	Delivered	Price
					(specify port) or	CIP	Duty	Duty Paid	delivered
					CIP named place	price	Unpaid	(DDP)	DDP
					(specify border	per	(DDU)		(col 4 x 8)
					point or place of	item			
					destination)	(col. 4			
						x 5)			
1	ULTRASONIC SURGICAL & ELECTROSURGICAL SYSTEM COMBINED IN 1 UNIT		1						
2	STRETCHER HYDRAULIC TYPE		3						
3	LARYNGOSCOPE, ADULT		7						
4	LARYNGOSCOPE SET		5						

 $^{^{3}\ \}mbox{If ADB, JICA}$ and WB funded projects, use IFB.

	FOR PEDIA							
[sig	gnature]			[i	n the capaci	ty of]		 ,
Dul	y authorized to sign Bi	d for an	d on be	half (of			

For Goods Offered From Within the Philippines

Name of Bidder	. Invitation to Bid ⁴	Number .	Page	of	

1	2	3		4	5	6	7	8	9	10
Lot	Description	Country of	Quantity	Unit	Unit	Transportation	Sales	Cost of	Total	Total Price
		origin			price	and Insurance	and	Incidental	Price,	delivered
					EXW	and all other	other	Services,	per unit	Final
					per	costs	taxes	if	(col	Destination
					item	incidental to	payable	applicable,	5+6+7+8)	(col 9) x
						delivery, per	if	per item		(col 4)
						item	Contract			
							is			
							awarded,			
							per item			
	ULTRASONIC SURGICAL &									
1	ELECTROSURGICAL SYSTEM		1							
	COMBINED IN 1 UNIT									
2	STRETCHER HYDRAULIC TYPE		3							
3	LARYNGOSCOPE, ADULT		7							
4	LARYNGOSCOPE SET FOR		5							
	PEDIA									

[signature]	[in the capacity of]
Duly authorized to sign Bid	for and on behalf of

 $^{^{4}\ \}mbox{If ADB, JICA}$ and WB funded projects, use IFB.

Technical Specifications

Statement of Compliance

Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause Error! Reference source not found. and/or GCC Clause Error! Reference source not found.

TECHNICAL SPECIFICATIONS Lot No. 01 ULTRASONIC SURGICAL & ELECTROSURGICAL SYSTEM COMBINED IN 1 UNIT Name of Manufacturer: Brand: ABC: Php 3,500,000.00 PURCHASER'S SPECIFICATION PURPOSE OF USE: Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: • Sealing and cutting simultaneously • Smart tissue sensing technology		Republic of the Philippines Department of Health				
Name of Manufacturer: Brand: ABC: Php 3,500,000.00 PURCHASER'S SPECIFICATION PURPOSE OF USE: Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously Country of Origin: Model: SUPPLIER'S SPECIFICATION Fill-in the suppliers specifications on the adjacent row of purchaser's specifications on the adjacent row of purchaser's specis.						
Brand: ABC: Php 3,500,000.00 PURCHASER'S SPECIFICATION PURPOSE OF USE: Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously SUPPLIER'S SPECIFICATION Fill-in the suppliers specifications on the adjacent row of purchaser's specs.	Lot No. 01 ULTRA	SONIC SURGICAL & ELECTROSURGICAL SYSTEM COMBINED IN 1 UNIT	(ty:1		
PURCHASER'S SPECIFICATION PURPOSE OF USE: Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously		r:	_		Or	rigin:
PURCHASER'S SPECIFICATION PURPOSE OF USE: Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously SUPPLIER'S SPECIFICATION Fill-in the suppliers specifications on the adjacent row of purchaser's specs.			N	Model:		
PURPOSE OF USE: Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously Technical service of the suppliers specifications on the adjacent row of purchaser's specs. Fill-in the suppliers specifications on the adjacent row of purchaser's specifications on the adjacent row of purchaser's specs.	ABC: Php 3,500,000.					
Ultrasonic Scalpel System provides surgeon with superior tissue dissection and vessel sealing capability with less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously		PURCHASER'S SPECIFICATION				
less thermal damage based on its smart tissue sensing technology. Adopt high-speed, high-precise acquisition and real-time feedback technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously specifications on the adjacent row of purchaser's specs.	PURPOSE OF USE:					
technology to automatically adjust the output voltage to ensure the stability and accuracy of the output power; Department: OR TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ◆ Sealing and cutting simultaneously purchaser's specs.				l		
TECHNICAL SPECIFICATION ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously	technology to automati					
ULTRASONIC GENERATOR SPECIFICATION: Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: ● Sealing and cutting simultaneously						
Voltage Input: 300 V or better Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: Sealing and cutting simultaneously	TECHNICAL SPECIFIC	<u>CATION</u>				
Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic Laparoscopic electrode & Bipolar coagulation forcep Operation mode: • Sealing and cutting simultaneously	-	ULTRASONIC GENERATOR SPECIFICATION:				
Laparoscopic electrode & Bipolar coagulation forcep Operation mode: Sealing and cutting simultaneously	_	Voltage Input: 300 V or better		·		
- Coperation mode: Operation mode: Sealing and cutting simultaneously □ Coperation mode: □ Coperatio	-	Parts: Ultrasonic scalpel, Electrosurgical pencil, Laparoscopic				
	-	Laparoscopic electrode & Bipolar coagulation forcep		·		
	-	Operation mode: • Sealing and cutting simultaneously			-	
The original regions of the state of the sta	-			· ·	-	
	-	0 0			-	
Power: ● Static (No-load) electric power: ≤20W		· , , , , , , , , , , , , , , , , , , ,	-			
Maximum electric power: ≤45W Connect 2-4 accessories to generator.			-			
• Index of power margin: >2		v v				

	Range of power level set: ● MAX power level ranges: 4 to 5	
	MIN power level ranges: 1 to 4	
	Tissue adaptive Algorithms : - Regulates energy delivery when needed for reducing	
	thermal damage; A tone change signals when tissue adaptive algorithms is running,	
	thereby energy deliveries more efficiently.	
Ultrasonic Scalpel:		
	System composition – Ultrasonic Scalpel	
	-Jaw, Waive Guide, Turning wheel, Activation button, Trigger & Handle	
	- Shears Length: 5cm, 9cm, 17cm, 14cm, 36cm (Lap Shears,	
	Thyroidectomy, Breast,: Open Abdomenal Shears, Gyne Surgery) -Transducer connector to Ultrasonic Generator	
	-Material : Titanium Alloy and latex free	
	-Shape: Curved Jaw design	
	-Maximum Sealing Vessel diameter: 7mm	
Electrosurgical	Flater vialance in the second	
	Electrosurgical pencil assembly	-
_	Electrode Handle: Disposable tip	-
10050005:=0	Electrode Handle: Resusable tip	-
ACCESSORIES:		
* Fully Cart or Tr	olly	
* Foot switch	Г. II. Т II.	
* Generator Cart	or Fully Trolly	
* 1 unit UPS with	AVR 2000VA	
* 1 unit UPS with Consumables / reage Suppliers Preferences	ents	
Consumables / reage	<u>ents</u>	
Consumables / reage Suppliers Preferences Spare parts per unit:	ents 6 units Fuses extra	
Consumables / reage Suppliers Preferences Spare parts per unit:	ents 6 units Fuses extra ATION AND UTILISATION	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi	ents 6 units Fuses extra ATION AND UTILISATION irements	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co	ents 6 units Fuses extra ATION AND UTILISATION irements	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co **** Supplier to perform ***Nurse staff/end-use	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning n installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning n installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installation	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning n installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALLA Pre-installation requi	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning n installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALLA Pre-installation requi Requirements for co	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform *** Surgeon / staff/en	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning in installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover id-user and biomed technician in coordination with the engineering department to affirm	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALLA Pre-installation requi Requirements for co *** Supplier to perform *** Surgeon / staff/en	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning in installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover id-user and biomed technician in coordination with the engineering department to affirm	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning in installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover id-user and biomed technician in coordination with the engineering department to affirm	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform *** Supplier	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning in installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover id-user and biomed technician in coordination with the engineering department to affirm	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform *** Supplier	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover end-user and biomed technician in coordination with the engineering department to affirm ion 5 Training of users in operation , basic and advance maintenance Including Fares, Meals shall be provided	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform *** Surgeon / staff/en completion of installati Training of users *** Atleast Two Weeks Accommodation and N *** Atleast 1 Surgeon	ATION AND UTILISATION irements mmissioning in installation, safety and operation checks before handover and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover installation, safety and operation checks before handover indicated and biomed technician in coordination with the engineering department to affirm ion Training of users in operation , basic and advance maintenance Including Fares, Meals shall be provided and 1 Biomed participants can attend for atleast One (1) Week Technical Training of	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform *** Surgeon / staff/en completion of installati Training of users *** Atleast Two Weeks Accommodation and N *** Atleast 1 Surgeon	ents 6 units Fuses extra ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover er and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover end-user and biomed technician in coordination with the engineering department to affirm ion 5 Training of users in operation , basic and advance maintenance Including Fares, Meals shall be provided	
Consumables / reage Suppliers Preferences Spare parts per unit: TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform ***Nurse staff/end-use completion of installati TRAINING, INSTALL Pre-installation requi Requirements for co *** Supplier to perform *** Supplier to perform *** Surgeon / staff/en completion of installati Training of users *** Atleast Two Weeks Accommodation and N **** Atleast 1 Surgeon Biomedical Technician	ATION AND UTILISATION irements mmissioning in installation, safety and operation checks before handover and biomed technician in coordination with the engineering department to affirm ion ATION AND UTILISATION irements mmissioning installation, safety and operation checks before handover installation, safety and operation checks before handover indicated and biomed technician in coordination with the engineering department to affirm ion Training of users in operation , basic and advance maintenance Including Fares, Meals shall be provided and 1 Biomed participants can attend for atleast One (1) Week Technical Training of	

2 years warranty should be provided by the supplier Parts = at least 3 years Maintenance and Service = at least 3 years Maintenance tasks *Semi-annual Preventive Maintenance/Calibration to be conducted by the company engineer in-charge *Advanced maintenance tasks required shall be documented *OR as the need arise Type of service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. **Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	Warranty		
Parts = at least 3 years Maintenance and Service = at least 3 years Maintenance tasks "Semi-annual Preventive Maintenance/Calibration to be conducted by the company engineer in-charge "Advanced maintenance tasks required shall be documented "OR as the need arise Type of service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	2 years warranty should be provided by the supplier		
Semi-annual Preventive Maintenance/Calibration to be conducted by the company engineer in-charge **Advanced maintenance tasks required shall be documented **OR as the need arise Type of service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided **DOCUMENTATION **Documentation requirements** - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	Parts = at least 3 years		
"Semi-annual Preventive Maintenance/Calibration to be conducted by the company engineer in-charge "Advanced maintenance tasks required shall be documented "OR as the need arise Type of service contract Service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
company engineer in-charge "Advanced maintenance tasks required shall be documented "OR as the need arise Type of service contract Service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	Maintenance tasks		
*Advanced maintenance tasks required shall be documented *OR as the need arise Type of service contract Service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	*Semi-annual Preventive Maintenance/Calibration to be conducted by the		
OR as the need arise Type of service contract Service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided **DOCUMENTATION **Documentation requirements** - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available **DECOMMISSIONING** Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. **Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Type of service contract Service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Service contract is recommended in case no in-house service experience is available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards		+	
available to ensure that preventive maintenance will be performed at regular intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
intervals. Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Pricing for service contracts should be negotiated before the system is purchased. Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	i i i		
Spare parts availability post-warranty: 10 years Software / Hardware upgrade availability: Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Software / Hardware upgrade availability : Routine software update provided DOCUMENTATION Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards		-	
- User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	DOCUMENTATION		
- User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	Documentation requirements		
- List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
- List of important spare parts and accessories, with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
- Trade-in with manufacturer if available DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
DECOMMISSIONING Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Estimated Life Span:10 years SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
SAFETY AND STANDARDS Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	DECOMMISSIONING		
Risk Classification Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	Estimated Life Span:10 years		
Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	SAFETY AND STANDARDS		
Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards	Risk Classification		
"Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards International standards			
Electrical safety conforms to standards International standards			
International standards	"Manufacturer / supplier should have ISO certificate for quality standard.		
	·		
LIFC / CE Conformity			
IEC / CE Conformity ISO			
EN			
Regional / Local Standards			
Regulations	Regulations		

	Republic of the Philippines Department of Health TECHNICAL SPECIFICATIONS							
Lot No. 02	Lot No. 02 STRETCHER HYDRAULIC TYPE Qty: 3							
Name of Man	ıfacturer:	Country of Origin:						
Brand:	Brand: Model:							
ABC: Php 75	0,000.00							

PURCHASER'S SPECIFICATION									PPLIER'S
							SF	PEC	CIFICATION
PURPOSE OF USE:									
Clinical or other Purpose: Stretcher allows for the transfer of patients easily and comfortably with a load capacity up to 750 mobility, and braking capabilities make the 7500 versatile for all your department needs such as C Transportation, Radiology or Cath Lab.	pounds. S DR, ED, P	upe ACL	rior J, Pa	stat atier	nt	spe adj	ecifi ace	icat ent	ne suppliers cions on the row of r's specs.
Department: Hospital				T	1				
Overview of Functional Requirements					L				
TECHNICAL CHARACTERISTICS									
Detailed Requirements:				T					
Specifications:			-		-		-		
<u>'</u>		Ī	-	F	-	ļ -	_		
Overall Length: 80-83" - Overall Width: 30-35"		-	-	H	-	-	-		
Height Range: 23" - 32.75"	-	-	-	Ħ	-	-	-		
Fowler Backrest: 0-90 degrees	-	+	H	H	 -	-	-		
Trendelenburg: +12 degrees -Reverse Trendelenburg: -7 degrees	-	1-	-	H	-	-	-		
Patient Surface: 75-80" x 30-35"	-	├	-	H	-	-	-		
Side Rails: 14-16" x 76-80"	- -	+-	-	片	-	-	-		
Casters: 8"	-	-	-	F	-	-	-		
		-	-	F	-	-	-		
Weight Capacity: 600-750 lbs		-	-	F	-	-	-		
Standard Features:	_ -	-	<u> -</u>	止	-	-	-		
Instant Steer Six Wheel Steering (IS6): The two IS6 wheels are centered directly under the chass maneuverability and control. Dual ball bearing wheels feature a softer wheel tread for improved gr	rip and tra	ease	ed 1.						
Rail System: Full length stainless steel side rails feature integrated push-pull handles with latch re all four corners of the stretcher. Easy one-handed operation. No hinges, mechanical connections debris or fluid. Easy to wipe clean and disinfect.					tra	р			
Oxygen Tank Holder: Ergonomic Quick-Release O2 Tank Holder is side mounted and convenient the patient surface to reduce caregiver bending.	ly located	und	er						
Center Column Support: Sturdy steel column with single hydraulic lift cylinder design uses sixteen bearing to smoothly raise and lower the patient.	precision	ball	-						
Hydraulic Height Adjustment Foot Pedals: Hydraulic Foot Pedals are located on each side of the scare zone. Color-coded pedals and function graphics are quickly and easily understood.	stretcher w	/ithir	n the	е ра	tier	nt			
Patient Lifts: The 7500 Series Stretcher has 5 7/8" H of clearance between floor and cowling to expatient lifts.	asily accor	nmc	date	e ar	nd u	ıtilize	Э		
Pneumatic Assist Fowler 600N (135lb.) force.									
Positive/negative Trendelenburg single pneumatic assist cylinders.									
Mechanical level lock									
Brake & Steer activated foot pedals									
Four corner IV pole receptacles				1					
Four restraint strap positions									
Full gray PVC perimeter bumper									
Latex-free				+					
Mattress is included				=					
Accessories per unit				t				ı	

Sterilization process for accessories	
Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification	
Should be FDA, CE or UL approved product.	
"Manufacturer / supplier should have ISO certificate for quality standard.	
Electrical safety conforms to standards for electrical safety	
Conforms to IEC-61010	
International standards: ISO Certified	
	•

	D	ublic of the Plepartment of IICAL SPECI	Health				
Lot No. 03	LARYNGOSCOPE, ADULT			Q	ty: 7		
				C	Country of Origin:		
Brand:				N	Model:		
ABC: Php 26	2,500.00						
PURCHASER'S SPECIFICATION					SUPPLIER'S SPECIFICATION		
PURPOSE OF	F USE: Clinical Purpose:					<u></u>	
Department:	<u>Hospital</u>			П	Fill-in the suppliers		
TECHNICAL	<u>CHARACTERISTICS</u>				specifications on the		
				Ħ	adjacent row of		
Stainless Steel Blades			П	purchaser's specs.			
Macintosh 3: /	Adult; Length: 130mm Width: 15mm	adult	Miller3: Adult L=155 W:14				
Macintosh 4:	Adult; Length: 155mm Width: 15mm	Large adult	Miller 4:Adult L=205 W: 14		adul t		
Complete Set with Carrying Case					1		
Fiber Optic G	reen Standard Laryngoscope LED Blade Re-L	Jsable				1	
Green Fiber C	Optic Stainless Steel Medical Grade (Blade fited	d on F/O Handl	e Only)				
Cold, Bright Light 6700 LUMINOUS.							
Autoclavable light guide tested to withstand over 600 cycles.					1		
1w LED lamp on the Handle has life of more than 1 Lakh Hrs.							
LED Lamp On the handle Provided White Bright Light.							
<u>Sterilization process for accessories</u> : Chemical sterilization- non corrosive chemical / dlsinfectant							
	s / reagents per unit Vhite LED Bulb AAA Batteries						

Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards : ISO Certified	

	Republic of the Philippines				
	Department of Health TECHNICAL SPECIFICATIONS				
	TECHNICAL SI ECHICATIONS				
Lot No. 04 LARYNGOSCOP	E SET FOR PEDIA		Qty:5		
Name of Manufacturer:		(Country of	Origin:	
Brand:		l l	Model:		
ABC: Php 187,500.00					
PU	JRCHASER'S SPECIFICATION			JPPLIER'S CIFICATION	
PURPOSE OF USE: Clinical Pu	rpose:				
Department: Hospital			Fill-in the suppliers specifications on the		
TECHNICAL CHARACTERISTIC	<u>s</u>				
			adjacent	I	
Stainless Steel Blades			purchase	er's specs.	
Macintosh 0 : Neonate Length: 77mm Width: 9mm	Miller 00 : Premature Length: 64mm Width: 11.5mm				
Macintosh 1 : Neonate Length: 90mm Width: 9mm	Miller0 : Neonate Length: 75mm Width: 11.5mm			,	
Macintosh 2 : Infant Length:	Miller1: Infant Length: 105mm Width:				
100mm Width: 12mm	11.5mm				
	Miller2: Child Length: 155mm Width:				
	14mm				
Complete Set with Carrying					
	Laryngoscope LED Blade Re-Usable				
Green Fiber Optic Stainless	Steel Medical Grade (Blade fited on F/O				

Handle Only)	
Cold, Bright Light 6700	1
LUMINOUS.	
Autoclavable light guide tested to withstand over 600 cycles.	
1w LED lamp on the Handle has life of more than 1 Lakh Hrs.	
LED Lamp On the handle Provided White Bright]
Light.	
1 Slim Handle & 1 Short	
Handle	
<u>Sterilization process for accessories</u> : Chemical sterilization- non corrosive chemical / dlsinfectant	
Consumables / reagents per unit	
2 pcs Spare White LED Bulb	
3 sets AA or AAA Batteries	
Regulatory Approval / Certification	1
Should be FDA, CE or UL approved product.	
"Manufacturer / supplier should have ISO certificate for quality standard.	
Electrical safety conforms to standards for electrical safety	
Conforms to IEC-61010	
International standards : ISO Certified]
WARRANTY: 1 YEAR	