



**ASTMMC - BIDS AND AWARDS COMMITTEE**

October 20, 2021

TO: **ALL CONCERN BIDDERS**  
 FROM: **ASTMMC – HBAC**

SUBJECT: **BID BULLETIN NO. 1**  
**Procurement of Mobile and Non-Mobile Infant Incubator with Phototherapy**  
**(IB No. 2021-11-03 (40))**

This Bid Bulletin No.1 is being issued to modify and to amend the following:

**1. INVITATION TO BID**

<b><u>PARTICULAR</u></b>	<b><u>CHANGES</u></b>
2. The [Adela Serra Ty Memorial Medical Center] now invites bids for the above Procurement Project. Delivery of the Goods is required by <b>[20 Calendar Days]</b> . Bidders should have completed, within [5 years] from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).	2. The [Adela Serra Ty Memorial Medical Center] now invites bids for the above Procurement Project. Delivery of the Goods is required by <b>[45 Calendar Days]</b> . Bidders should have completed, within [5 years] from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).

**2. SCHEDULE OF REQUIREMENTS**

<b><u>PARTICULAR</u></b>	<b><u>CHANGES</u></b>
Delivered, Weeks/Months	<b>Delivered, Weeks/Months</b>
20 Calendar Days	<b>45 Calendar Days</b>

**3. TECHNICAL SPECIFICATIONS**

<b>Item No. 1 – MOBILE INFANT INCUBATOR WITH PHOTOTHERAPY</b>	
<b><u>PARTICULAR</u></b>	<b><u>CHANGES</u></b>
<b>Standard Configurations</b> Main Body (Including the hood controller bassinet, Internal Battery, and Observation Lamp)	<b>Standard Configurations</b> Main Body (Including the hood controller bassinet, Internal Battery, and <b>Observation Lamp –separate or not integrated is acceptable</b> )



<b>Spareparts per unit</b>  (2)Two sets of spare fuses per units (if non-resettable fuses used).  (2) sets of Filter	<b>Spareparts per unit</b>  ((2)Two sets of spare fuses per units(if non-resettable fuses used). (2) sets of Filter <b>2pcs of Fuses</b> <b>1 pcs Disposable Skin temperature Sensor</b>
--	---

<b>Item No. 2 - INFANT INCUBATOR WITH PHOTOTHERAPY</b>	
<b><u>PARTICULAR</u></b>	<b><u>CHANGES</u></b>
<b><u>TECHNICAL CHARACTERISTICS</u></b>  AC Power and DC Power Alternatively DC 12V Or 24V Power on the Ambulance	<b><u>TECHNICAL CHARACTERISTICS</u></b>  AC Power and DC Power Alternatively DC 12V <b>Or 24V Power on the Ambulance- Deleted</b>
<b>Spareparts per unit</b>  (2)Two sets of spare fuses per units (if non-resettable fuses used).  (2) sets of Filter	<b>Spareparts per unit</b>  ((2)Two sets of spare fuses per units(if non-resettable fuses used). (2) sets of Filter <b>2pcs of Fuses</b> <b>1 pcs Disposable Skin temperature Sensor</b>

**Note:**

1. Single Largest Completed Contract (SLCC) should be similar of the item to be bid.
2. Bid Proposal should be in four (4) copies- 1 original and 3 copies.
3. ASTMMC is now accepting online payment of Bidding Documents thru our Landbank Account

Account Name: ASTMMC-TF  
 Account No.: 0792-0109-54

Proof of payment or deposit slip should be sent to: [astmmcprocurement@gmail.com](mailto:astmmcprocurement@gmail.com)

4. Bid Proposals maybe sent via courier and should be received by the Bids and Awards Committee on or before November 03, 2021; 10:00am. Online submission of bid proposal is not allowed.
5. Net Financial Contracting Capacity should be NOTARIZED.

For further information you may coordinate with the Bids and Awards Secretariat at Tel. No. 086-211-4306 or you can email your concerns at [astmmcprocurement@gmail.com](mailto:astmmcprocurement@gmail.com)

Thank You,

  
**VOLTAIRE S. EGNORA, MD, FPCP**  
 HBAC – Chairman