



REQUEST FOR QUOTATION (RFQ)

For Procurement of **Various Medical Equipment**

RFQ No. 048

1. The Adela Serra Ty Memorial Medical Center, through the GOP-GAA intends to apply the sum of **Eight Hundred Seventy Nine Thousand Pesos Only (Php 879, 000.00)** being the approved Budget for the Contract (ABC) to payments for the Procurement of **VARIOUS MEDICAL EQUIPMENT** for which this Request for Quotation is issued.
2. The ASTMMC, hereinafter referred to as “the Purchaser” now request you to submit price quotation for the supply and delivery of the following:

Item No.	Description	Qty.	Unit
1	SPHYGMOMANOMETER PORTABLE ANEROID (ADULT)	15	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Detailed Requirements: Characteristics Type:aneroid Description Product Name : Clock Type Aneroid Sphygmomanometer, Mobile Description : Large 6" gauge with easily read character Mobile Features easy to move Specification : Mobile Aneroid Sphyg. with coiled tubing & inflation system Equipped with latex or non-latex (PVC or neoprene) bulb		
	Accessories per unit 1 adult Cuff with Bladder & 1 adult Cuff with Bladder 1 extra Valve Pump with pressure release		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dIsinfectant		
	Consumables / reagents		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Laboratory staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	SAFETY AND STANDARDS		
	Risk Classification		
	Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards		
	Sterilization process for accessories		



	Chemical sterilization- non corrosive chemical / disinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards ISO Certified		
2	SPHYGMOMANOMETER PORTABLE ANEROID (PEDIA)	15	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Detailed Requirements: Characteristics Type:aneroid Description Product Name : Clock Type Aneroid Sphygmomanometer, Mobile Description : Large 6" gauge with easily read character Mobile Features easy to move Specification : Mobile Aneroid Sphyg. with coiled tubing & inflation system Equipped with latex or non-latex (PVC or neoprene) bulb		
	Accessories per unit 1 adult Cuff with Bladder & 1 adult Cuff with Bladder 1 extra Valve Pump with pressure release		
	Consumables / reagents		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Laboratory staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	Risk Classification		
	Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards ISO Certified		
3	STETHOSCOPE (ADULT)	4	UNIT



	PURPOSE OF USE:		
	Technical Features:		
	Dual Head Stethoscope		
	High acoustic sensitivity		
	Monitor and assess a wide range of patients		
	Detect normal and abnormal sounds and rhythms		
	Tunable diaphragms on both the adult and pediatric sides of the chestpiece		
	Single-piece, easy to clean diaphragm		
	Pediatric side converts to a traditional open bell by replacing the single-piece diaphragm with a non-chill sleeve		
	Next-generation tubing with improved resistance to skin oils and alcohol for longer tubing life, less likely to pick up stains		
	Chestpiece is precision-machined into an aesthetically pleasing, less angular shape		
	Stem features an open side indicator		
	Useful in non-critical care environments like medical offices, OB/GYN, ambulatory clinic and urgent care		
	5 year warranty		
	Tubing is 27" in length		
	Spare parts per unit : 1 sets Earpiece 1 sets Bell adult and Neonate		
	DOCUMENTATION		
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Trade-in with manufacturer if available		
	WARRANTY AND MAINTENANCE		
	Warranty : 6months warranty		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	International standards ISO Certified		
4	STETHOSCOPE (PEDIA)	4	UNIT
	PURPOSE OF USE:		
	Technical Features:		
	Dual Head Stethoscope		
	High acoustic sensitivity		
	Monitor and assess a wide range of patients		
	Detect normal and abnormal sounds and rhythms		
	Tunable diaphragms on both the adult and pediatric sides of the chestpiece		
	Single-piece, easy to clean diaphragm		
	Pediatric side converts to a traditional open bell by replacing the single-piece diaphragm with a non-chill sleeve		



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Next-generation tubing with improved resistance to skin oils and alcohol for longer tubing life, less likely to pick up stains		
	Chestpiece is precision-machined into an aesthetically pleasing, less angular shape		
	Stem features an open side indicator		
	Useful in non-critical care environments like medical offices, OB/GYN, ambulatory clinic and urgent care		
	5 year warranty		
	Tubing is 27" in length		
	Spare parts per unit : 1 sets Earpiece 1 sets Bell adult and Neonate		
	DOCUMENTATION		
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Trade-in with manufacturer if available		
	WARRANTY AND MAINTENANCE		
	Warranty : 6months warranty		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	International standards ISO Certified		
5	PORTABLE HEAVY DUTY SUCTION MACHINE WITH ACCESSORIES	40	UNIT
	PURPOSE OF USE		
	Clinical or other Purpose: Drive Medical Suction Machine is a lightweight, suction pump used to remove secretions from the mouth and throat. This heavy duty suction machine can be used in a professional setting, or for home use. The Heavy Duty Suction Pump is a professional medical suctioning device that produces a vacuum of 0 to 560 mmHg. A 800 ml collection bottle (suction canister) and 6 foot connecting tubing is included.		
	Clinical Department/Ward		
	Hospitalization		
	Overview of Functional Requirements:		
	Features New anti-bacteria filter design helps prevent contamination. High quality vacuum regulator and anti-vibration vacuum gauge setting. Oil-Less, permanent lubrication. Collection bottle with over-flow protection. Superior performance and high reliability for continuous operation design. Easy to carry and clean		
	TECHNICAL CHARACTERISTICS		
	Weight: 5.5 kg (12.1 lbs) or better Maximum Vacuum: 560 +/- mm Hg.		



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Flow Range: 40 LPM. Sound Level: 58 dBA Electrical Requirements: AC 220 V - 60Hz Power Consumption: 70 watts Bacteria Filter pore size: 0.027 microns Collection Bottle Capacity: 800 ml Operating Temperature: 32° ~ 104° F (0° ~ 40° C) Operating Humidity: 95% RH		
	PHYSICAL/CHEMICAL CHARACTERISTICS		
	Motor pump must Metal hard Structure		
	ACCESSORIES, CONSUMABLES, SPARE PARTS, OTHER COMPONENTS		
	Accessories		
	Disposable Suction Canister Kit includes 10" tubing, elbow connector, 72" suction tubing, and a hydrophobic filter.		
	Consumables / reagents		
	20 pcs Hydrophobic Filters per unit 1 Tubing, and 1 Kits per unit		
	Spare parts		
	(2)Two sets of spare fuses Medical units select them according to their needs, ensuring compatibility with the brand and model of the equipment.		
	TRAINING, INSTALLATION AND UTILISATION		
	Pre-installation requirements		
	Training of users *** Training of users in operation and basic maintenance and calibration shall be provided *** Supplier to issue Certificate of Training *** Advanced maintenance tasks required shall be documented		
	User care *** Unit layout to enable easy cleaning and sterilization of all surfaces.		
	WARRANTY AND MAINTENANCE		
	Warranty : 2 years warranty		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	SAFETY AND STANDARDS		
	Risk Classification		
	Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard.		



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
6	PULSE OXIMETER (ADULT) PORTABLE FINGERTIPS	12	Unit
	PURPOSE OF USE:		
	Pulse oximetry is a noninvasive and painless test that measures your oxygen saturation level, or the oxygen levels in your blood		
	Technical Features:		
	100% Brand New and High Quality		
	Reliable Accuracy And Durability		
	2-Direction Display Adjustable, Allow You To View Your Results In Different Direction		
	Low Voltage Indicator, Real-Time Spot-Checks		
	Low Power Consumption, 50 Hours Continuous To Work More Than 8 Seconds		
	Self-Adjusting Finger Clamp Plus Simple One-Button Design Allows For Easy Operation		
	Comes With A Neck/Wrist Cord, Convenient For Nurse And Doctor To Hang On The Neck		
	Small Portable Size Makes It Easy To Handle And Carry		
	Suitable For Daily Use, Measure Your Health At Any Time		
	The Finger Pulse Oximeter Is An Affordable And Accurate Way To Check Pulse Rates And Blood Oxygen Saturation Levels		
	Widely Used In Hospital, Home Healthcare, Oxygen Bar, Community Medical Centre, Alpine Area, Sports Healthcare etc.		
	DOCUMENTATION		
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided with Sticker in Equipment as proof of calibration - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available		
	WARRANTY AND MAINTENANCE		
	Warranty : 6months warranty		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	SAFETY AND STANDARDS		
	Risk Classification		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dIsinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product.		



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	"Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
7	PULSE OXIMETER (NEONATE) PORTABLE FINGERTIPS	12	UNIT
	PURPOSE OF USE:		
	Pulse oximetry is a noninvasive and painless test that measures your oxygen saturation level, or the oxygen levels in your blood		
	Technical Features:		
	100% Brand New and High Quality		
	Reliable Accuracy And Durability		
	2-Direction Display Adjustable, Allow You To View Your Results In Different Direction		
	Low Voltage Indicator, Real-Time Spot-Checks		
	Low Power Consumption, 50 Hours Continuous To Work More Than 8 Seconds		
	Self-Adjusting Finger Clamp Plus Simple One-Button Design Allows For Easy Operation		
	Comes With A Neck/Wrist Cord, Convenient For Nurse And Doctor To Hang On The Neck		
	Small Portable Size Makes It Easy To Handle And Carry		
	Suitable For Daily Use, Measure Your Health At Any Time		
	The Finger Pulse Oximeter Is An Affordable And Accurate Way To Check Pulse Rates And Blood Oxygen Saturation Levels		
	Widely Used In Hospital, Home Healthcare, Oxygen Bar, Community Medical Centre, Alpine Area, Sports Healthcare etc.		
	with Battery		
	DOCUMENTATION		
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided with Sticker in Equipment as proof of calibration - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available		
	WARRANTY AND MAINTENANCE		
	Warranty : 6months warranty		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	SAFETY AND STANDARDS		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard.		



	Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
8	OXYGEN GAUGE WITH HUMIDIFIER	12	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Detailed Requirements: Features: Product Specification Regulator Type High Pressure Regulator Rated flow 0.2, 0.4, 0.6 Automation Grade Semi-Automatic Phase Single phase		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant		
	WARRANTY AND MAINTENANCE		
	Warranty : Maintenance and Service = 6mons		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010 Certificate of Calibration		
10	INFRARED THERMOMETER	3	UNIT
	PURPOSE OF USE: Clinical or other Purpose: digital thermometer can perform non-contact measurements of surface temperatures. The most attractive features of the digital thermometer are its clear laser point, the wide temperature range and the adjustable emissivity (depending on the material).		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Detailed Requirements:		
	Measuring range -50 ... +1050 °C		
	Resolution 0.1 °C		
	Accuracy ±4 °C (between -50 ... 0 °C)		
	±3 % of reading ± 2 °C (between 0 ... +1050 °C)		
	Response time <250 ms		
	Distance-to-spot ratio 30:1		
	Adjustable emissivity 0,1 ... 1,0 (depending on the material)		
	Spectral range 8 ... 14 µm		
	Laser pointer single laser point; extra bright; for aiming		
	Laser class Class 2; <1 mW		
	Functions Max-Min measurements		
	continuous measurement		
	illuminated display		
	Standards EMC; EN 50081-1, 50082-1, 60825-1		
	Power supply 1 x 9 V battery		



	Operating conditions 0 ... +50 °C / 10 ... 95 % RH		
	Storing conditions -20 ... +60 °C / < 80 % RH		
	Dimensions 155 x 98 x 27 mm		
	Weight 171 g		
	Accessories per unit : Battery 2 Sets		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards ISO Certified		
9	PATSLIDE PATIENT TRANSFER	8	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Made from thick semi-flexible thermoplastic material		
	12 longitudinal ridges on base to reduce sideway movement of the board.		
	Cut out on both sides to assist with transfers and carrying.		
	Static Insulation		
	Radio Translucent		
	Impact Resistant		
	Accessories per unit		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards: ISO Certified		
10	EMERGENCY MEDICAL RING CUTTER & JEWELRY REMOVER	1	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Finger Ring Cutter Features		
	Chrome finish		
	6 inches in length		
	Sliding safety lever		
	Protective plate		
	Razor sharp serrated saw		
	Sold by the each		
	The safety plate slides between finger and ring, allowing you to cut rings safely!		
	Accessories per unit		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant		
	Regulatory Approval / Certification		



	Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards: ISO Certified		
11	TREATMENT TROLLEY	6	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	Nursing Trolley		
	Stainless steel		
	Two shelves with railing		
	Size: atleast W670 * D450mm* H900		
	With S.S. basin, drawer, dustbin		
	Noiseless casters with brake		
	Packing Size: atleast 700*480*800mm(1 pcs /carton)		
	Gross Weight:17kg		
	Net Weight:16kg		
	Accessories per unit		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards: ISO Certified		
12	DRESSING CART	6	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department: Hospital		
	TECHNICAL CHARACTERISTICS		
	4 units 4" caster (2 with brake) With tray, bowl and trash can Dressing fitting (option)		
	Chrome plated materials		
	With Soaking materials		
	Built in drawers		
	With movable wheels		
	Included one pail bucket		
	Accessories per unit		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant		
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010		
	International standards: ISO Certified		
13	NEBULIZER MACHINE	8	UNIT



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	PURPOSE OF USE: Clinical or other Purpose:		
	Department Hospital		
	TECHNICAL CHARACTERISTICS		
	Particulars: Dimensions		
	Size: atleast 10.1" D x 10.5" H x 6.5" W		
	Electrical requirements: 1.3 amps, 220VAC 60Hz		
	Sound level: 51-55 dBA		
	Power consumption: 90 watts maximum		
	Maximum compressor pressure: 35 psig		
	Compressor free air flow: 12 lpm		
	Operating temperature range: +40° to +104°F (+5 to +40°C)		
	Operating Humidity: up to 95% non-condensing		
	Storage/Transport Temperature Range: -40° to +158°F (-40° to +70°C)		
	Storage/Transport Humidity: up to 95% non-condensing		
	Storage/Transport Pressure: 7.3 to 15.4 psia		
	Supplied Nebulizer Nebulization Rate: 0.15 ml/min or greater		
	Supplied Nebulizer MMAD: 5 microns or less		
	Supplied nebulizer capacity: 10 ml (cc)		
	Accessories per unit		
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / d/lsinfectant		
	Consumables / reagents ***20pcs FILTER per UNIT		
	Sterilization process for accessories Ease of sterilization thru commercially available/common d/lsinfectant		
	User care Casing to be splash proof and cleanable with alcohol or chlorine wipes		
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover ***End-user and biomed technician in coordination with the engineering department to affirm completion of installation		
	Warranty : Maintenance and Service = 1 year		
	International standards ISO Certified		
	Regulations		
14	GOOSE NECK LAMP WITH COVER	6	UNIT
	PURPOSE OF USE: Clinical or other Purpose:		
	Department Hospital		
	Detailed Requirements:		
	Stainless Stand		
	Locally Made		
	Consumables / reagents ***5pcs Extra Bulb 100W		
	International standards ISO Certified		
	Regulations		



Republic of the Philippines
Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



3. A set of technical specifications for the above items are provided in Attachment 1. All items listed under the purchasers specifications must be complied on a pass-fail basis. Failure to meet any one of the requirements will result to rejection.
4. Bidding procedures will be conducted in accordance with the provisions of the Implementing Rules and Regulations-A (IRR-A) of Republic Act 9184.
5. It is the intent of the Purchaser to evaluate the quotation for each item separately, and award will be made to the quotation or combination of quotations resulting in the overall lowest cost. Quotation for one or more or all items will be considered. Contract Award will be made to the lowest evaluated quotation (s) meeting purchaser's technical specifications.
6. Quotations must be delivered at the HBAC Office, ASTMMC, Tandag, Surigao del Sur not later than 4:00 pm. Of **February 9, 2021**
7. Your prices must be quoted in Philippine Peso and must include the unit price and total price, inclusive of all taxes to be paid and other incidental cost to the delivery site if the contract is awarded.
8. All quotations must be typewritten and must be placed in sealed envelope marked "PROCUREMENT OF **VARIOUS MEDICAL EQUIPMENT (RFQ No. 048)**"
9. Quotations shall be valid for sixty (60) calendar days from the Opening of Bids.
10. The delivery period shall be within fifteen (15) calendar days from receipt of the Notice to Proceed (NTP).
11. DELIVERY SITE: Supply Office, Adela Serra Ty Memorial Medical Center, Capitol Hills, Tandag City.
12. The applicable rate for late deliveries is one tenth (1/10) of one percent (1%) of the cost of the unperformed portion for every day of delay. The maximum deduction shall be ten percent (10%) of the amount of contract. Once the cumulative amount of liquidated damages reaches ten percent (10%) of the amount of the contract, the procuring entity shall rescind the contract, without prejudice to other courses of action and remedies open to it.
13. The ASTMMC reserves the right to accept or reject any quotation, and to annul the bidding process and reject all quotations at any time prior to Contract award, without thereby incurring any liability to the affected bidder or bidders and to waive any minor defects or infirmities therein and to accept such quotation as may be considered advantageous to the government.



Republic of the Philippines
Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



14. The prospective bidder shall submit the following:
- Quotation Form
 - Technical Specification
 - Certificate of PHIL-GEPS Registration

VOLTAIRE S. EGNORA, MD, FPCP

HBAC Chairperson

ASTMMC

Fax Number: (083) 2114306

E-mail: astmmc_hbac@yahoo.com.ph



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



QUOTATION FORM

Received By: _____
Signature over Printed Name
Date & Time: _____

Date: _____

RFQ No. 048

The Chairman

The Bids and Award Committee on Goods

ASTMMC

- 1) Having examined the subject Request for Quotation including the Technical Specifications, we, the undersigned offer to supply and deliver the following items (with our unit price and total)

Item No.	Description	Qty.	Unit	Unit Price	Total Price
1	SPHYGMOMANOMETER PORTABLE ANEROID (ADULT)	15	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Detailed Requirements: Characteristics Type:aneroid Description Product Name : Clock Type Aneroid Sphygmomanometer, Mobile Description : Large 6" gauge with easily read character Mobile Features easy to move Specification : Mobile Aneroid Sphyg. with coiled tubing & inflation system Equipped with latex or non-latex (PVC or neoprene) bulb				
	Accessories per unit 1 adult Culf with Bladder & 1 adult Culf with Bladder 1 extra Valve Pump with pressure release				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant				
	Consumables / reagents				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Laboratory staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	SAFETY AND STANDARDS				
	Risk Classification				
	Regulatory Approval / Certification Should be a CE or UL approved product.				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	"Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
	International standards ISO Certified				
2	SPHYGMOMANOMETER PORTABLE ANEROID (PEDIA)	15	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Detailed Requirements: Characteristics Type:aneroid Description Product Name : Clock Type Aneroid Sphygmomanometer, Mobile Description : Large 6" gauge with easily read character Mobile Features easy to move Specification : Mobile Aneroid Sphyg. with coiled tubing & inflation system Equipped with latex or non-latex (PVC or neoprene) bulb				
	Accessories per unit 1 adult Culf with Bladder & 1 adult Culf with Bladder 1 extra Valve Pump with pressure release				
	Consumables / reagents				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Laboratory staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	Risk Classification				
	Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant				
	Regulatory Approval / Certification				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
	International standards ISO Certified				
3	STETHOSCOPE (ADULT)	4	UNIT		
	PURPOSE OF USE:				
	Technical Features:				
	Dual Head Stethoscope				
	High acoustic sensitivity				
	Monitor and assess a wide range of patients				
	Detect normal and abnormal sounds and rhythms				
	Tunable diaphragms on both the adult and pediatric sides of the chestpiece				
	Single-piece, easy to clean diaphragm				
	Pediatric side converts to a traditional open bell by replacing the single-piece diaphragm with a non-chill sleeve				
	Next-generation tubing with improved resistance to skin oils and alcohol for longer tubing life, less likely to pick up stains				
	Chestpiece is precision-machined into an aesthetically pleasing, less angular shape				
	Stem features an open side indicator				
	Useful in non-critical care environments like medical offices, OB/GYN, ambulatory clinic and urgent care				
	5 year warranty				
	Tubing is 27" in length				
	Spare parts per unit : 1 sets Earpiece 1 sets Bell adult and Neonate				
	DOCUMENTATION				
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Trade-in with manufacturer if available				
	WARRANTY AND MAINTENANCE				
	Warranty : 6months warranty				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	International standards ISO Certified				
4	STETHOSCOPE (PEDIA)	4	UNIT		
	PURPOSE OF USE:				
	Technical Features:				
	Dual Head Stethoscope				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	High acoustic sensitivity				
	Monitor and assess a wide range of patients				
	Detect normal and abnormal sounds and rhythms				
	Tunable diaphragms on both the adult and pediatric sides of the chestpiece				
	Single-piece, easy to clean diaphragm				
	Pediatric side converts to a traditional open bell by replacing the single-piece diaphragm with a non-chill sleeve				
	Next-generation tubing with improved resistance to skin oils and alcohol for longer tubing life, less likely to pick up stains				
	Chestpiece is precision-machined into an aesthetically pleasing, less angular shape				
	Stem features an open side indicator				
	Useful in non-critical care environments like medical offices, OB/GYN, ambulatory clinic and urgent care				
	5 year warranty				
	Tubing is 27" in length				
	Spare parts per unit : 1 sets Earpiece 1 sets Bell adult and Neonate				
	DOCUMENTATION				
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Trade-in with manufacturer if available				
	WARRANTY AND MAINTENANCE				
	Warranty : 6months warranty				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	International standards ISO Certified				
5	PORTABLE HEAVY DUTY SUCTION MACHINE WITH ACCESSORIES	40	UNIT		
	PURPOSE OF USE				
	Clinical or other Purpose: Drive Medical Suction Machine is a lightweight, suction pump used to remove secretions from the mouth and throat. This heavy duty suction machine can be used in a professional setting, or for home use. The Heavy Duty Suction Pump is a professional medical suctioning device that produces a vacuum of 0 to 560 mmHg. A 800 ml collection bottle (suction canister) and 6 foot connecting tubing is included.				
	Clinical Department/Ward				
	Hospitalization				
	Overview of Functional Requirements:				
	Features				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	<p>New anti-bacteria filter design helps prevent contamination. High quality vacuum regulator and anti-vibration vacuum gauge setting. Oil-Less, permanent lubrication. Collection bottle with over-flow protection. Superior performance and high reliability for continuous operation design. Easy to carry and clean</p>				
	TECHNICAL CHARACTERISTICS				
	<p>Weight: 5.5 kg (12.1 lbs) or better Maximum Vacuum: 560 +/- mm Hg. Flow Range: 40 LPM. Sound Level: 58 dBA Electrical Requirements: AC 220 V - 60Hz Power Consumption: 70 watts Bacteria Filter pore size: 0.027 microns Collection Bottle Capacity: 800 ml Operating Temperature: 32° ~ 104° F (0° ~ 40° C) Operating Humidity: 95% RH</p>				
	PHYSICAL/CHEMICAL CHARACTERISTICS				
	Motor pump must Metal hard Structure				
	ACCESSORIES, CONSUMABLES, SPARE PARTS, OTHER COMPONENTS				
	Accessories				
	Disposable Suction Canister Kit includes 10" tubing, elbow connector, 72" suction tubing, and a hydrophobic filter.				
	Consumables / reagents				
	20 pcs Hydrophobic Filters per unit 1 Tubing, and 1 Kits per unit				
	Spare parts				
	(2)Two sets of spare fuses Medical units select them according to their needs, ensuring compatibility with the brand and model of the equipment.				
	TRAINING, INSTALLATION AND UTILISATION				
	Pre-installation requirements				
	<p>Training of users *** Training of users in operation and basic maintenance and calibration shall be provided *** Supplier to issue Certificate of Training *** Advanced maintenance tasks required shall be documented</p>				
	<p>User care *** Unit layout to enable easy cleaning and sterilization of all surfaces.</p>				
	WARRANTY AND MAINTENANCE				
	Warranty : 2 years warranty				
	<p>Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation</p>				
	SAFETY AND STANDARDS				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Risk Classification				
	Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
6	PULSE OXIMETER (ADULT) PORTABLE FINGERTIPS	12	Unit		
	PURPOSE OF USE:				
	Pulse oximetry is a noninvasive and painless test that measures your oxygen saturation level, or the oxygen levels in your blood				
	Technical Features:				
	100% Brand New and High Quality				
	Reliable Accuracy And Durability				
	2-Direction Display Adjustable, Allow You To View Your Results In Different Direction				
	Low Voltage Indicator, Real-Time Spot-Checks				
	Low Power Consumption, 50 Hours Continuous To Work More Than 8 Seconds				
	Self-Adjusting Finger Clamp Plus Simple One-Button Design Allows For Easy Operation				
	Comes With A Neck/Wrist Cord, Convenient For Nurse And Doctor To Hang On The Neck				
	Small Portable Size Makes It Easy To Handle And Carry				
	Suitable For Daily Use, Measure Your Health At Any Time				
	The Finger Pulse Oximeter Is An Affordable And Accurate Way To Check Pulse Rates And Blood Oxygen Saturation Levels				
	Widely Used In Hospital, Home Healthcare, Oxygen Bar, Community Medical Centre, Alpine Area, Sports Healthcare etc.				
	DOCUMENTATION				
	Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided with Sticker in Equipment as proof of calibration - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	- List of important spare parts and accessories with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available				
	WARRANTY AND MAINTENANCE				
	Warranty : 6months warranty				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	SAFETY AND STANDARDS				
	Risk Classification				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
7	PULSE OXIMETER (NEONATE) PORTABLE FINGERTIPS	12	UNIT		
	PURPOSE OF USE:				
	Pulse oximetry is a noninvasive and painless test that measures your oxygen saturation level, or the oxygen levels in your blood				
	Technical Features:				
	100% Brand New and High Quality				
	Reliable Accuracy And Durability				
	2-Direction Display Adjustable, Allow You To View Your Results In Different Direction				
	Low Voltage Indicator, Real-Time Spot-Checks				
	Low Power Consumption, 50 Hours Continuous To Work More Than 8 Seconds				
	Self-Adjusting Finger Clamp Plus Simple One-Button Design Allows For Easy Operation				
	Comes With A Neck/Wrist Cord, Convenient For Nurse And Doctor To Hang On The Neck				
	Small Portable Size Makes It Easy To Handle And Carry				
	Suitable For Daily Use, Measure Your Health At Any Time				
	The Finger Pulse Oximeter Is An Affordable And Accurate Way To Check Pulse Rates And Blood Oxygen Saturation Levels				
	Widely Used In Hospital, Home Healthcare, Oxygen Bar, Community Medical Centre, Alpine Area, Sports Healthcare etc.				
	with Battery				
	DOCUMENTATION				
	Documentation requirements				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	<ul style="list-style-type: none"> - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided with Sticker in Equipment as proof of calibration - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available 				
	WARRANTY AND MAINTENANCE				
	Warranty : 6months warranty				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	SAFETY AND STANDARDS				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
8	OXYGEN GAUGE WITH HUMIDIFIER	12	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Detailed Requirements: Features: Product Specification Regulator Type High Pressure Regulator Rated flow 0.2, 0.4, 0.6 Automation Grade Semi-Automatic Phase Single phase				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant				
	WARRANTY AND MAINTENANCE				
	Warranty : Maintenance and Service = 6mons				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010 Certificate of Calibration				
10	INFRARED THERMOMETER	3	UNIT		
	PURPOSE OF USE:				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	<p>Clinical or other Purpose: digital thermometer can perform non-contact measurements of surface temperatures. The most attractive features of the digital thermometer are its clear laser point, the wide temperature range and the adjustable emissivity (depending on the material).</p>				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Detailed Requirements:				
	Measuring range -50 ... +1050 °C				
	Resolution 0.1 °C				
	Accuracy ±4 °C (between -50 ... 0 °C)				
	±3 % of reading ± 2 °C (between 0 ... +1050 °C)				
	Response time <250 ms				
	Distance-to-spot ratio 30:1				
	Adjustable emissivity 0,1 ... 1,0 (depending on the material)				
	Spectral range 8 ... 14 µm				
	Laser pointer single laser point; extra bright; for aiming				
	Laser class Class 2; <1 mW				
	Functions Max-Min measurements				
	continuous measurement				
	illuminated display				
	Standards EMC; EN 50081-1, 50082-1, 60825-1				
	Power supply 1 x 9 V battery				
	Operating conditions 0 ... +50 °C / 10 ... 95 % RH				
	Storing conditions -20 ... +60 °C / < 80 % RH				
	Dimensions 155 x 98 x 27 mm				
	Weight 171 g				
	Accessories per unit : Battery 2 Sets				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
	International standards ISO Certified				
9	PATSLIDE PATIENT TRANSFER	8	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Made from thick semi-flexible thermoplastic material				
	12 longitudinal ridges on base to reduce sideway movement of the board.				
	Cut out on both sides to assist with transfers and carrying.				
	Static Insulation				
	Radio Translucent				
	Impact Resistant				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Accessories per unit				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
10	INTERNATIONAL STANDARDS: ISO CERTIFIED EMERGENCY MEDICAL RING CUTTER & JEWELRY REMOVER	1	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Finger Ring Cutter Features				
	Chrome finish				
	6 inches in length				
	Sliding safety lever				
	Protective plate				
	Razor sharp serrated saw				
	Sold by the each				
	The safety plate slides between finger and ring, allowing you to cut rings safely!				
	Accessories per unit				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
	INTERNATIONAL STANDARDS: ISO CERTIFIED				
11	TREATMENT TROLLEY	6	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	Nursing Trolley				
	Stainless steel				
	Two shelves with railing				
	Size: atleast W670 * D450mm* H900				
	With S.S. basin, drawer, dustbin				
	Noiseless casters with brake				
	Packing Size: atleast 700*480*800mm(1 pcs /carton)				
	Gross Weight:17kg				
	Net Weight:16kg				
	Accessories per unit				
	Sterilization process for accessories				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Chemical sterilization- non corrosive chemical / disinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
	International standards: ISO Certified				
12	DRESSING CART	6	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department: Hospital				
	TECHNICAL CHARACTERISTICS				
	4 units 4" caster (2 with brake) With tray, bowl and trash can Dressing fitting (option)				
	Chrome plated materials				
	With Soaking materials				
	Built in drawers				
	With movable wheels				
	Included one pail bucket				
	Accessories per unit				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant				
	Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010				
	International standards: ISO Certified				
13	NEBULIZER MACHINE	8	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department Hospital				
	TECHNICAL CHARACTERISTICS				
	Particulars: Dimensions				
	Size: atleast 10.1" D x 10.5" H x 6.5" W				
	Electrical requirements: 1.3 amps, 220VAC 60Hz				
	Sound level: 51-55 dBA				
	Power consumption: 90 watts maximum				
	Maximum compressor pressure: 35 psig				
	Compressor free air flow: 12 lpm				
	Operating temperature range: +40° to +104°F (+5 to +40°C)				
	Operating Humidity: up to 95% non-condensing				
	Storage/Transport Temperature Range: -40° to +158°F (-40° to +70°C)				
	Storage/Transport Humidity: up to 95% non-condensing				
	Storage/Transport Pressure: 7.3 to 15.4 psia				
	Supplied Nebulizer Nebulization Rate: 0.15 ml/min or greater				



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



	Supplied Nebulizer MMAD: 5 microns or less				
	Supplied nebulizer capacity: 10 ml (cc)				
	Accessories per unit				
	Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant				
	Consumables / reagents ***20pcs FILTER per UNIT				
	Sterilization process for accessories Ease of sterilization thru commercially available/common dlsinfectant				
	User care Casing to be splash proof and cleanable with alcohol or chlorine wipes				
	Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** End-user and biomed technician in coordination with the engineering department to affirm completion of installation				
	Warranty : Maintenance and Service = 1 year				
	International standards ISO Certified				
	Regulations				
14	GOOSE NECK LAMP WITH COVER	6	UNIT		
	PURPOSE OF USE: Clinical or other Purpose:				
	Department Hospital				
	Detailed Requirements:				
	Stainless Stand				
	Locally Made				
	Consumables / reagents ***5pcs Extra Bulb 100W				
	International standards ISO Certified				
	Regulations				
TOTAL QUOTATION					

Note:

- **Brochure on the quoted item should be submitted**
- **Quotations must be forwarded on or before 4:00pm of Feb. 9, 2021**

2. We undertake, if our Quotation or Bid is accepted, to deliver the above goods within the fifteen (15) days delivery period.
3. We agree to abide by this Quotation/Bid for a period of sixty (60) days after the dated deadline of submission specified in your RFQ.
4. We understand that payment for items delivered will be made to the winning supplier within thirty (30) days after inspection and acceptance of goods delivered.

Company Name : _____

Address: _____

Signature Over Printed Name: _____



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Republic of the Philippines Adela Serra Ty Memorial Medical Center TECHNICAL SPECIFICATIONS	
Purchaser's Specifications	Supplier's Specifications (please state "comply" or "not comply")
SPHYGMOMANOMETER PORTABLE ANEROID (ADULT)	
PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
Detailed Requirements: Characteristics Type: aneroid Description Product Name : Clock Type Aneroid Sphygmomanometer, Mobile Description : Large 6" gauge with easily read character Mobile Features easy to move Specification : Mobile Aneroid Sphyg. with coiled tubing & inflation system Equipped with latex or non-latex (PVC or neoprene) bulb	
Accessories per unit 1 adult Cuff with Bladder & 1 adult Cuff with Bladder 1 extra Valve Pump with pressure release	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Consumables / reagents	
Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Laboratory staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation	
SAFETY AND STANDARDS	
Risk Classification	
Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



ISO Certified	
SPHYGMOMANOMETER PORTABLE ANEROID (PEDIA)	
PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
Detailed Requirements: Characteristics Type: aneroid Description Product Name : Clock Type Aneroid Sphygmomanometer, Mobile Description : Large 6" gauge with easily read character Mobile Features easy to move Specification : Mobile Aneroid Sphyg. with coiled tubing & inflation system Equipped with latex or non-latex (PVC or neoprene) bulb	
Accessories per unit 1 adult Culf with Bladder & 1 adult Culf with Bladder 1 extra Valve Pump with pressure release	
Consumables / reagents	
Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Laboratory staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation	
Risk Classification	
Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards ISO Certified	
STETHOSCOPE (ADULT)	
PURPOSE OF USE:	
Technical Features:	
Dual Head Stethoscope	
High acoustic sensitivity	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Monitor and assess a wide range of patients	
Detect normal and abnormal sounds and rhythms	
Tunable diaphragms on both the adult and pediatric sides of the chestpiece	
Single-piece, easy to clean diaphragm	
Pediatric side converts to a traditional open bell by replacing the single-piece diaphragm with a non-chill sleeve	
Next-generation tubing with improved resistance to skin oils and alcohol for longer tubing life, less likely to pick up stains	
Chestpiece is precision-machined into an aesthetically pleasing, less angular shape	
Stem features an open side indicator	
Useful in non-critical care environments like medical offices, OB/GYN, ambulatory clinic and urgent care	
5 year warranty	
Tubing is 27" in length	
Spare parts per unit : 1 sets Earpiece 1 sets Bell adult and Neonate	
DOCUMENTATION	
Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Trade-in with manufacturer if available	
WARRANTY AND MAINTENANCE	
Warranty : 6months warranty	
Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation	
International standards ISO Certified	
STETHOSCOPE (PEDIA)	
PURPOSE OF USE:	
Technical Features:	
Dual Head Stethoscope	
High acoustic sensitivity	
Monitor and assess a wide range of patients	
Detect normal and abnormal sounds and rhythms	
Tunable diaphragms on both the adult and pediatric sides of the chestpiece	
Single-piece, easy to clean diaphragm	
Pediatric side converts to a traditional open bell by replacing the single-piece diaphragm with a non-chill sleeve	
Next-generation tubing with improved resistance to skin oils and alcohol for longer tubing life, less likely to pick up stains	
Chestpiece is precision-machined into an aesthetically pleasing, less angular shape	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Stem features an open side indicator	
Useful in non-critical care environments like medical offices, OB/GYN, ambulatory clinic and urgent care	
5 year warranty	
Tubing is 27" in length	
Spare parts per unit : 1 sets Earpiece 1 sets Bell adult and Neonate	
DOCUMENTATION	
Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Trade-in with manufacturer if available	
WARRANTY AND MAINTENANCE	
Warranty : 6months warranty	
Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation	
International standards ISO Certified	
PORTABLE HEAVY DUTY SUCTION MACHINE WITH ACCESSORIES	
PURPOSE OF USE	
Clinical or other Purpose: Drive Medical Suction Machine is a lightweight, suction pump used to remove secretions from the mouth and throat. This heavy duty suction machine can be used in a professional setting, or for home use. The Heavy Duty Suction Pump is a professional medical suctioning device that produces a vacuum of 0 to 560 mmHg. A 800 ml collection bottle (suction canister) and 6 foot connecting tubing is included.	
Clinical Department/Ward	
Hospitalization	
Overview of Functional Requirements:	
Features New anti-bacteria filter design helps prevent contamination. High quality vacuum regulator and anti-vibration vacuum gauge setting. Oil-Less, permanent lubrication. Collection bottle with over-flow protection. Superior performance and high reliability for continuous operation design. Easy to carry and clean	
TECHNICAL CHARACTERISTICS	
Weight: 5.5 kg (12.1 lbs) or better Maximum Vacuum: 560 +/- mm Hg. Flow Range: 40 LPM. Sound Level: 58 dBA Electrical Requirements: AC 220 V - 60Hz	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Power Consumption: 70 watts Bacteria Filter pore size: 0.027 microns Collection Bottle Capacity: 800 ml Operating Temperature: 32° ~ 104° F (0° ~ 40° C) Operating Humidity: 95% RH	
PHYSICAL/CHEMICAL CHARACTERISTICS	
Motor pump must Metal hard Structure	
ACCESSORIES, CONSUMABLES, SPARE PARTS, OTHER COMPONENTS	
Accessories	
Disposable Suction Canister Kit includes 10" tubing, elbow connector, 72" suction tubing, and a hydrophobic filter.	
Consumables / reagents	
20 pcs Hydrophobic Filters per unit 1 Tubing, and 1 Kits per unit	
Spare parts	
(2)Two sets of spare fuses Medical units select them according to their needs, ensuring compatibility with the brand and model of the equipment.	
TRAINING, INSTALLATION AND UTILISATION	
Pre-installation requirements	
Training of users *** Training of users in operation and basic maintenance and calibration shall be provided *** Supplier to issue Certificate of Training *** Advanced maintenance tasks required shall be documented	
User care *** Unit layout to enable easy cleaning and sterilization of all surfaces.	
WARRANTY AND MAINTENANCE	
Warranty : 2 years warranty	
Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation	
SAFETY AND STANDARDS	
Risk Classification	
Regulatory Approval / Certification Should be a CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Conforms to IEC-61010	
PULSE OXIMETER (ADULT) PORTABLE FINGERTIPS	
PURPOSE OF USE:	
Pulse oximetry is a noninvasive and painless test that measures your oxygen saturation level, or the oxygen levels in your blood	
Technical Features:	
100% Brand New and High Quality	
Reliable Accuracy And Durability	
2-Direction Display Adjustable, Allow You To View Your Results In Different Direction	
Low Voltage Indicator, Real-Time Spot-Checks	
Low Power Consumption, 50 Hours Continuous To Work More Than 8 Seconds	
Self-Adjusting Finger Clamp Plus Simple One-Button Design Allows For Easy Operation	
Comes With A Neck/Wrist Cord, Convenient For Nurse And Doctor To Hang On The Neck	
Small Portable Size Makes It Easy To Handle And Carry	
Suitable For Daily Use, Measure Your Health At Any Time	
The Finger Pulse Oximeter Is An Affordable And Accurate Way To Check Pulse Rates And Blood Oxygen Saturation Levels	
Widely Used In Hospital, Home Healthcare, Oxygen Bar, Community Medical Centre, Alpine Area, Sports Healthcare etc.	
DOCUMENTATION	
Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided with Sticker in Equipment as proof of calibration - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available	
WARRANTY AND MAINTENANCE	
Warranty : 6months warranty	
Requirements for commissioning *** Supplier to perform installation, safety and operation checks before handover *** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation	
SAFETY AND STANDARDS	
Risk Classification	
Sterilization process for accessories	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
PULSE OXIMETER (NEONATE) PORTABLE FINGERTIPS	
PURPOSE OF USE:	
Pulse oximetry is a noninvasive and painless test that measures your oxygen saturation level, or the oxygen levels in your blood	
Technical Features:	
100% Brand New and High Quality	
Reliable Accuracy And Durability	
2-Direction Display Adjustable, Allow You To View Your Results In Different Direction	
Low Voltage Indicator, Real-Time Spot-Checks	
Low Power Consumption, 50 Hours Continuous To Work More Than 8 Seconds	
Self-Adjusting Finger Clamp Plus Simple One-Button Design Allows For Easy Operation	
Comes With A Neck/Wrist Cord, Convenient For Nurse And Doctor To Hang On The Neck	
Small Portable Size Makes It Easy To Handle And Carry	
Suitable For Daily Use, Measure Your Health At Any Time	
The Finger Pulse Oximeter Is An Affordable And Accurate Way To Check Pulse Rates And Blood Oxygen Saturation Levels	
Widely Used In Hospital, Home Healthcare, Oxygen Bar, Community Medical Centre, Alpine Area, Sports Healthcare etc.	
with Battery	
DOCUMENTATION	
Documentation requirements - User, technical and maintenance manuals to be supplied in English language - at least 3 copies - Certificate of calibration and inspection to be provided with Sticker in Equipment as proof of calibration - List of equipment and procedures required for local calibration and routine maintenance to be provided by the supplier - List of important spare parts and accessories with their part numbers and cost to be provided by the supplier - Trade-in with manufacturer if available	
WARRANTY AND MAINTENANCE	
Warranty : 6months warranty	
Requirements for commissioning	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



<p>*** Supplier to perform installation, safety and operation checks before handover</p> <p>*** Staff/end-user and biomed technician in coordination with the engineering department to affirm completion of installation</p>	
SAFETY AND STANDARDS	
<p>Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010</p>	
OXYGEN GAUGE WITH HUMIDIFIER	
PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
<p>Detailed Requirements: Features: Product Specification Regulator Type High Pressure Regulator Rated flow 0.2, 0.4, 0.6 Automation Grade Semi-Automatic Phase Single phase</p>	
<p>Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant</p>	
WARRANTY AND MAINTENANCE	
Warranty : Maintenance and Service = 6mons	
<p>Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010 Certificate of Calibration</p>	
INFRARED THERMOMETER	
<p>PURPOSE OF USE: Clinical or other Purpose: digital thermometer can perform non-contact measurements of surface temperatures. The most attractive features of the digital thermometer are its clear laser point, the wide temperature range and the adjustable emissivity (depending on the material).</p>	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
Detailed Requirements:	
Measuring range -50 ... +1050 °C	
Resolution 0.1 °C	
Accuracy ±4 °C (between -50 ... 0 °C)	
±3 % of reading ± 2 °C (between 0 ... +1050 °C)	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



Response time <250 ms	
Distance-to-spot ratio 30:1	
Adjustable emissivity 0,1 ... 1,0 (depending on the material)	
Spectral range 8 ... 14 µm	
Laser pointer single laser point; extra bright; for aiming	
Laser class Class 2; <1 mW	
Functions Max-Min measurements	
continuous measurement	
illuminated display	
Standards EMC; EN 50081-1, 50082-1, 60825-1	
Power supply 1 x 9 V battery	
Operating conditions 0 ... +50 °C / 10 ... 95 % RH	
Storing conditions -20 ... +60 °C / < 80 % RH	
Dimensions 155 x 98 x 27 mm	
Weight 171 g	
Accessories per unit : Battery 2 Sets	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards ISO Certified	
PATSLIDE PATIENT TRANSFER	
PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
Made from thick semi-flexible thermoplastic material	
12 longitudinal ridges on base to reduce sideway movement of the board.	
Cut out on both sides to assist with transfers and carrying.	
Static Insulation	
Radio Translucent	
Impact Resistant	
Accessories per unit	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards: ISO Certified	
EMERGENCY MEDICAL RING CUTTER & JEWELRY REMOVER	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
Finger Ring Cutter Features	
Chrome finish	
6 inches in length	
Sliding safety lever	
Protective plate	
Razor sharp serrated saw	
Sold by the each	
The safety plate slides between finger and ring, allowing you to cut rings safely!	
Accessories per unit	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards: ISO Certified	
TREATMENT TROLLEY	
PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	
Nursing Trolley	
Stainless steel	
Two shelves with railing	
Size: atleast W670 * D450mm* H900	
With S.S. basin, drawer, dustbin	
Noiseless casters with brake	
Packing Size: atleast 700*480*800mm(1 pcs /carton)	
Gross Weight:17kg	
Net Weight:16kg	
Accessories per unit	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / disinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards: ISO Certified	
DRESSING CART	
PURPOSE OF USE: Clinical or other Purpose:	
Department: Hospital	
TECHNICAL CHARACTERISTICS	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



4 units 4" caster (2 with brake) With tray, bowl and trash can Dressing fitting (option)	
Chrome plated materials	
With Soaking materials	
Built in drawers	
With movable wheels	
Included one pail bucket	
Accessories per unit	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant	
Regulatory Approval / Certification Should be FDA, CE or UL approved product. "Manufacturer / supplier should have ISO certificate for quality standard. Electrical safety conforms to standards for electrical safety Conforms to IEC-61010	
International standards: ISO Certified	
NEBULIZER MACHINE	
PURPOSE OF USE: Clinical or other Purpose:	
Department Hospital	
TECHNICAL CHARACTERISTICS	
Particulars: Dimensions	
Size: atleast 10.1" D x 10.5" H x 6.5" W	
Electrical requirements: 1.3 amps, 220VAC 60Hz	
Sound level: 51-55 dBA	
Power consumption: 90 watts maximum	
Maximum compressor pressure: 35 psig	
Compressor free air flow: 12 lpm	
Operating temperature range: +40° to +104°F (+5 to +40°C)	
Operating Humidity: up to 95% non-condensing	
Storage/Transport Temperature Range: -40° to +158°F (-40° to +70°C)	
Storage/Transport Humidity: up to 95% non-condensing	
Storage/Transport Pressure: 7.3 to 15.4 psia	
Supplied Nebulizer Nebulization Rate: 0.15 ml/min or greater	
Supplied Nebulizer MMAD: 5 microns or less	
Supplied nebulizer capacity: 10 ml (cc)	
Accessories per unit	
Sterilization process for accessories Chemical sterilization- non corrosive chemical / dlsinfectant	
Consumables / reagents ***20pcs FILTER per UNIT	
Sterilization process for accessories Ease of sterilization thru commercially available/common dlsinfectant	
User care Casing to be splash proof and cleanable with alcohol or chlorine wipes	
Requirements for commissioning	



Republic of the Philippines
 Department of Health
ADELA SERRA TY MEMORIAL MEDICAL CENTER
 Capitol Hills, Telaje, Tandag City, Surigao del Sur
Telefax No. 086-211-4306
E-mail: astmmc_hbac@yahoo.com.ph
Facebook Account: AstmmcProbac



*** Supplier to perform installation, safety and operation checks before handover ***End-user and biomed technician in coordination with the engineering department to affirm completion of installation	
Warranty : Maintenance and Service = 1 year	
International standards ISO Certified	
Regulations	
GOOSE NECK LAMP WITH COVER	
PURPOSE OF USE: Clinical or other Purpose:	
Department Hospital	
Detailed Requirements:	
Stainless Stand	
Locally Made	
Consumables / reagents ***5pcs Extra Bulb 100W	
International standards ISO Certified	
Regulations	